The pilot of new requirement on procurement of Medical Devices (MD) by the Department of Health



## Purpose

- ▶ To enhance the customer protection for Department of Health (DH) as a buyer of medical devices
- ► To ensure the safety, quality and performance of the medical devices procured and to ensure the availability of after-sale (post-market) support from suppliers
- ► To strike a balance between value for money purchase while enhancing public safety
- ► To raise public awareness of the importance of MD safety and pave way for the introduction of long-term statutory control

#### Procurement Requirements

- MD being procured by DH is <u>preferably</u> be listed under the Medical Device Administrative Control System (MDACS)
- Quotation: A preference will be given in the quotation for listed MDs under the MDACS
  - If there are more than one lowest conforming offers offering the same price, preference will be given to the MD offered which are listed under the MDACS.
- ► Tender: A technical score will be given in the marking scheme for listed MDs under the MDACS
  - In assessment criterion under the tender evaluation, marks will be given for MD offered which are preferably be listed under the MDACS.
- ▶ If the proposed MD is accepted by the Government, the Tenderer shall be bound by such offer and shall be deemed to form part of the Contract.
- For details of the procurement requirements for a particular MD procurement, please refer to the tender documents.

# Submissions to be provided by Supplier for indicating the compliance of the desirable requirements

- To indicate in the schedule of tender/ quotation whether the Item(s) offered is listed under MDACS
- To provide HKMD listing number and a valid copy of the Certificate of Listing under MDACS.
  - ► For application for listing of Medical Device, please refer to MDD website (<a href="https://www.mdd.gov.hk/en/mdacs/listing-application/medical-device/index.html">https://www.mdd.gov.hk/en/mdacs/listing-application/medical-device/index.html</a>)
  - ► The vetting and approval of an application for listing a device should normally be completed within 12 weeks following the submission of the application and all the required supporting information, including labelling samples. (ref.: Guidance Notes GN-01)

## Implementation of the Pilot

- ► To allow stakeholders to get familiarized with the new procurement requirement, a pilot of the initiative has been implemented w.e.f 1st Oct 2022.
- Under the pilot, certain MD being procured by the DH will include the new procurement requirement in the quotation/tender exercises that the MD under purchase is <u>preferably</u> be listed under the MDACS.
- Subject to the experience gained from the pilot, DH will extend to include more MD into the scheme.
- Traders are encouraged to submit the listing application for their MDs under MDACS as soon as possible.

#### Online Medical Device Classification Tools

► To facilitate traders and public users to determine if a product is a MD, and if so, its associated class. On-line tools have been developed on website <a href="https://www.mdd.gov.hk/en/mdacs/online-tools/index.html">https://www.mdd.gov.hk/en/mdacs/online-tools/index.html</a>

Online Tools

(For Medical Device Administrative Control System Only)

- ► <u>Is Your Product A Medical Device?</u>
- ► General Medical Device Classification Program
- ► In Vitro Diagnostic Medical Device Classification Program

## **Enquiries**

- For general enquiries on the new procurement requirement and application for listing of MDs under MDACS:
  - Medical Device Division, Department of Health

► Tel: 3107 8484

Fax: 3157 1286

► Email: mdd@dh.gov.hk

Website: https://www.mdd.gov.hk/

- For enquiries on the other requirements of a particular MD procurement tender/ quotation:
  - ▶ Please contact the officer(s) of the procurement service as stipulated in the quotation/tender document.